


RIDA qLine[®] autoBlot Fully automated processor for allergy diagnostics



The RIDA qLine[®] autoBlot in combination with the RIDA qLine[®] Soft is a fully automated processor for the RIDA qLine[®] Allergy membranes



All parts can easily be removed for cleaning purposes except the barcode reader and the orbital shaker.



How to work with the RIDA qLine® autoBlot

The RIDA qLine[®] autoBlot must be connected to a separate computer on which the RIDA qLine[®] Soft is installed over a network. The RIDA qLine[®] Soft is the centerpiece of the workstation responsible for creating the worklist and the evaluation of the measurements.

Up to 4 processors can be connected to the software and operated in parallel.

1. Connected to a LIS

This device shows its high professionality and user friendliness when connected to a Laboratory Information System (LIS).

Procedure


After entering up to 36 barcoded sample tubes randomly into the sample carousel of the RIDA qLine[®] autoBlot and pressing the start button, the detected barcodes are transferred to the RIDA qLine[®] Soft which calls for the requests to each barcode at the LIS, creates a worklist from it and sends it to the machine.



Immediately after receiving the worklist the run can be prepared by the user. The software shows the volumes of the reagents needed 1 and guides the user through all steps for preparing the run 2. The allocation of the membranes according to the worklist is also displayed 3 in order to make the assembly of the tray with the membranes easy and safe.

Worklists The reagent volumes required for this run are					Worklists Complete each of the listed tasks to continue		3 Worklists Please load the pends into the tray, as specified below	
Wash Buffer Conjugate DI Water	85.5 ml 4 ml 4 ml		4 ml		Coad Samples	Load Tray Load Tip Rack Load Washing Buffer	Provide Gale (in proceeding on the large set substance) concerned In the set of the large set substance of the large set set set substance of the large set set set set set set set set set se	
C C C C C C C C C C C C C C C C C C C			•	7.540	Empty Waste Bottle	Empty Tip Waste		


When the run is started the software shows the remaining time, the current time or the completion time 4. The desired display can be chosen by the user.





2. Not connected to a LIS

If the user is not working with a LIS the RIDA qLine[®] autoBlot can also be used in combination with the RIDA qLine[®] Soft. The user first has to enter manually the tests which should be run into the RIDA qLine[®] Soft. Then the run can be started as described above.



Consumables

The pipet tips and the glass vials for the reagents can be ordered at R-Biopharm:

- Pipet tips (Art. No. Z0013)
- Glass vials (Art. No. Z0011)
- Screw caps (Art. No. Z0012)

Technical data

Instrument type	Fully automated bench top	Fuses	5A/250V UL approved –
			Ø 5 x 20 mm
Processing capability	1 to 36 samples per run		
Processing time	Typically 3.5 hours for	Dimensions	600 (H) x 620 (D) x 780 (W) mm
	36 samples	Weight	65 kg
Temperature control	PID under software control	Consumption value (energy)	400 W (max)
Processing volumes	10 μL to 1000 μL	Acoustic noise level (dB)	~ 60 dB
Dispensing mode & accuracy	Piston pump/peristaltic pumps:	Electromagnetic emissions	Class A 30 MHz to 1 GHz
	+/- 10 %	Radiated field immunity	10 V/m 80 MHz to 1 GHz 3 V/m 1.4 GHz to 2 GHz
Software	Embedded software		
Software			
Voltage & frequency	Electrical power source		1 V/m 2 GHz to 2.7 GHz
i onage of nequency	100 to 240 V AC 10A 50/60 Hz		

Barcodes

JAN/UPC/EAN incl. add on, Codabar NW-7, Code 11, Code 39, Code 93, Code 128, GS1-128 (EAN-128), GS1 DataBar (RSS), IATA, Industrial 2of5, Interleaved 2of5, ISBN-ISMN-ISSN, Matrix 2of5, MSI/ Plessey, S-Code, Telepen, Tri-Optic, UK/Plessey Postal code: Chinese Post, Korean Postal Authority



For further information please visit: www.r-biopharm.com