



## PRODUCT INFORMATION

**Product Name:**

PROCEEDx Influenza B and Respiratory Syncytial Virus (HRSV) Positive Sample

**Catalogue Number:**

VP-14-02

**Product Description:**

This product contains inactivated influenza B and respiratory syncytial virus in storage buffer.

## SPECIFICATIONS

**Strains:**

*Influenza B/Hong Kong/5/72. ICTV 00.046.0.04.001*

*Human respiratory syncytial virus, Long strain. ICTV 01.048.2.01.001*

**Storage:**

Store this preparation at 2 °C to 8 °C.

**Shelf Life:**

72 weeks from manufacturing date.

**Volume:**

1 mL

## USE

**Instructions For Use:**

Use PROCEEDx product as per sample handling instructions specified by device or instrument manufacturer.

**Intended Use:**

The product is to be used **for immunoassay and/or nucleic acid test** instrument (device) verification/validation, manufacturing QMS, educational and R&D purposes.

**NOT FOR USE IN DIAGNOSTIC PROCEDURES. FOR RESEARCH USE ONLY.**

## ASSISTANCE

If you have any questions regarding the production, testing or use of this antigen, please send them together with any relevant data to Microbix Technical Services, at [customer.service@microbix.com](mailto:customer.service@microbix.com) or fax 905-361-8911. Your complete satisfaction with the performance of this product is important to us.

## HAZARDS

We are aware of no specific biological hazards associated with this product. The preparation has been inactivated and should contain no infectious material however it should be handled as if potentially infectious. Generally accepted good laboratory practices appropriate to biological reagents should be employed. The buffer solution contains 0.02% sodium azide (CAS No. 26628-22-8) as a preservative.






## SPECIAL HANDLING

No special handling required for this product.

# QUALITY CONTROL INFORMATION

Test Kit	Target	Result
Quickvue Influenza A & B Catalogue # 20183IN	Influenza A & B	<b>Influenza B Positive</b> <b>Influenza A Negative</b>
OSOM Influenza A & B Catalogue # 190	Influenza A & B	<b>Influenza B Positive</b> <b>Influenza A Negative</b>
Binax NOW Influenza A & B Catalogue # 416-000	Influenza A & B	<b>Influenza B Positive</b> <b>Influenza A Negative</b>
Xpect RSV Catalogue # R24601	Respiratory Syncytial Virus	<b>RSV Positive</b>
SAS RSV Catalogue # 041130	Respiratory Syncytial Virus	<b>RSV Positive</b>
Binax NOW RSV Catalogue # 430-022	Respiratory Syncytial Virus	<b>RSV Positive</b>
SAS Adenovirus Catalogue # 067020	Adenovirus	<b>Adenovirus Negative</b>
SAS Rotavirus Catalogue # 069020	Rotavirus	<b>Rotavirus Negative</b>
Quickvue Chlamydia Catalogue # OB006	Chlamydia	<b>Chlamydia Negative</b>

qPCR Test Kit	Analyte	Result
Xpert FLU/RSV Cepheid	Influenza A	<b>Not Detected</b>
Xpert FLU/RSV Cepheid	Influenza B	<b>Detected</b>
Xpert FLU/RSV Cepheid	Respiratory Syncytial Virus	<b>Detected</b>

	Catalog Number
	Lot Number
	Temperature Limitation
	Expiration Date
	Biological Risk

