



### PRODUCT INFORMATION

**Product Name:**

PROCEEDx Influenza B Positive Swab

**Catalogue Number:**

VP-S-14-01

**Product Description:**

This product contains dry swab inactivated Influenza B virus.

### SPECIFICATIONS

**Strain:**

Influenza B/Hong Kong/5/72. ICTV 00.046.0.04.001

**Storage:**

Store this preparation at 2 °C to 25 °C.

**Shelf Life:**

1 year from manufacturing date.

### USE

**Instructions For Use:**

Use PROCEEDx swab product as per sample handling instructions specified by device or instrument manufacturer.

**Intended Use:**

The product is to be used **for immunoassay and/or nucleic acid test** instrument (device) verification/validation, manufacturing QMS, educational and R&D purposes.

**NOT FOR USE IN DIAGNOSTIC PROCEDURES. FOR RESEARCH USE ONLY.**

### ASSITANCE

If you have any questions regarding the production, testing or use of this antigen, please send them together with any relevant data to Microbix Technical Services, at [customer.service@microbix.com](mailto:customer.service@microbix.com) or fax 905-361-8911. Your complete satisfaction with the performance of this product is important to us.

### HAZARDS

We are aware of no specific biological hazards associated with this product. The preparation has been inactivated and should contain no infectious material however it should be handled as if potentially infectious. Generally accepted good laboratory practices appropriate to biological reagents should be employed.




### SPECIAL HANDLING

No special handling required for this product.

## QUALITY CONTROL INFORMATION

Test / Kit	Target	Result
OSOM Flu A & B Catalogue # 190	Influenza A & B	<b>Influenza B</b> <b>Positive</b> <b>Influenza A</b> <b>Negative</b>

qPCR Test Kit	Analyte	Result
Sekisui Silaris Flu A/B kit	Influenza A&B	<b>Influenza B</b> <b>Detected</b> <b>Influenza A</b> <b>Not Detected</b>

<b>REF</b>	Catalog Number
<b>LOT</b>	Lot Number
	Temperature Limitation
	Expiration Date
	Biological Risk

