

RIDASCREEN® TDM Wash buffer 20x

REF 3096P00-VK



1. Intended use

For *in vitro* diagnostic use. The RIDASCREEN® TDM Wash buffer 20x (3096P00-VK) is intended as a wash buffer for use with the RIDASCREEN® IFX Monitoring (G09041), RIDASCREEN® Anti-IFX Antibodies (G09042), RIDASCREEN® ADM Monitoring (G09043), RIDASCREEN® Anti-ADM Antibodies (G09044), RIDASCREEN® VDZ Monitoring (G09045), RIDASCREEN® GLM Monitoring (G09047), and RIDASCREEN® UST Monitoring (G09049) test.

The product is intended for professional use.

2. Summary and explanation of the test

The RIDASCREEN® TDM Wash buffer 20x is offered as an accessory for the RIDASCREEN® tests (see Section 1. Intended use).

3. Test principle

The RIDASCREEN® TDM Wash buffer 20x is concentrated 20-fold and must be used as described in the instructions for use for the RIDASCREEN® tests (see Section 1. Intended use).

The RIDASCREEN® TDM Wash buffer 20x is used for washing the RIDASCREEN® microtiter plate.

4. Reagents provided

Table 1: Reagents provided for the RIDASCREEN® TDM Wash buffer 20x

Kit components	Amount		Description
TDM Wash buffer 20x <i>Transparent lid</i>	1 x	50 mL	Wash buffer, detergent in phosphate buffer solution and antimicrobial agents (concentrated 20-fold)

5. Storage instructions

Please follow the handling guidelines in Table 2 and store the product directly after use according to the information specified. After the expiration date has passed or the recommended storage period of the opened product has elapsed, the quality guarantee is no longer valid.

The RIDASCREEN® **TDM Wash buffer 20x** is supplied as a 20-fold concentrate and must be diluted with distilled or deionized water 1:20 before use (add 50 mL concentrate to a 1000 mL graduated cylinder and fill with distilled water to 1000 mL).

Table 2: Storage conditions and information

	Storage temperature	Maximum storage time	Additional notes on storage
unopened	2 - 8 °C	Can be used until the printed expiration date	-
diluted wash buffer	2 - 8 °C	≤ 4 weeks	-

6. Reagents required but not provided

6.1 Reagents

Reagents
RIDASCREEN® Test (see Section 1. Intended use)

6.2 Laboratory equipment

The laboratory equipment needed for the RIDASCREEN® test is required for performing the test (see Section 1. Intended use).

7. Warnings and precautions for the users

Only qualified laboratory personnel may use this product.

The guidelines for working in medical laboratories (good laboratory practice) must be followed. Adhere strictly to the operating manual when performing the RIDASCREEN® tests. Do not pipette samples or reagents using your mouth. Avoid contact with broken skin and mucous membranes. Wear personal protective equipment (appropriate gloves, lab coat, safety glasses) when handling reagents and samples, and wash hands after finishing using the product. Do not smoke, eat, or drink in areas where samples are handled.

Hazardous materials are indicated according to labeling obligations.

Further details on the Safety Data Sheet (SDS) can be found under the item number at <https://clinical.r-biopharm.com/search/>.

Users are responsible for the proper disposal of the RIDASCREEN®

TDM Wash buffer 20x after use. For disposal, please adhere to national regulations.

For users in the European Union: Report all serious adverse events associated with the product to R-Biopharm AG and the relevant national authorities.

8. Test procedure

Before use, bring the RIDASCREEN® **TDM Wash buffer 20x** to room temperature (20 - 25 °C). The RIDASCREEN® **TDM Wash buffer 20x** must be used as described in the instructions for use for the RIDASCREEN® tests (see Section 1. Intended use).

9. Limitations of the method










1. The RIDASCREEN® **TDM Wash buffer 20x** is intended for use only with said RIDASCREEN® TDM tests (see Section 1. Intended use).
2. This product should be used in compliance with the regulation on good laboratory practice (GLP). When using the RIDASCREEN® **TDM Wash buffer 20x**, operators must precisely follow the manufacturer's instructions for the RIDASCREEN® test.

10. Version history


Version number	Section and designation
2021-12-09	Release version

11. Explanation of symbols

General symbols

	For <i>in vitro</i> diagnostic use
	Observe operating manual
	Batch number
	Use before
	Storage temperature
	Item number
	Number of tests
	Date of manufacture
	Manufacturer

Test-specific symbols

	Wash buffer (concentrated 20-fold)
---	------------------------------------